

IMPORTANT INFORMATION TO READ and RETURN

Installation Requirements for a Whitley A25 Anaerobic Workstation

Thank you for choosing one of our products for your laboratory. To enable our engineers to perform an efficient, trouble-free installation please study, complete and fax this form to us on **01274 531197**. Should you have any questions, please do not hesitate to contact us, as we are here to help. When we have received the completed form, our Service Department will contact you to arrange a mutually convenient installation date.

The following information represents the ideal requirement.

Please contact us IMMEDIATELY if your intended location does not match this specification.

				▼
Access Rec	quirements			
For access, th	ne dimensions b	elow should be	taken into account when checking the size of doorways, lifts, stairs, etc	
Space Requ	uirements			
The weight of	the equipment	is 74kg. If bencl	h mounted, the bench allocated must be flat, level and of sufficient size to support the bas	e fully.
External Di	mensions*			
Width	Depth	Height		
mm	mm	mm		
790	720	710		
* Please Note In addition to		s stated above, a	allow a localised protrusion of 90mm at rear to accommodate condenser bottle, bracket,	

Gas Requirements

of 500mm at the right hand side of the unit is required for user/service access.

The incoming gas supplies must be terminated near the right hand side of the main chamber and fitted with leak-proof taps and pressure gauges.

gas and electrical supplies. If benchmounted, a further minimum clearance of 500mm is required above the unit and a minimum clearance

The gas lines to which the equipment is attached are the responsibility of the user and should be constructed, tested and maintained to the standards specified within the British Compressed Gasses Association (BCGA) Code of Practice CP4 (or international equivalent). Gas lines previously used for flammable gases must be purged prior to re-use.

Regulators should be fitted in accordance with the information contained in the table below and the various pressures strictly adhered to.

Gas Type	Connection Details	Cylinder Regulator Required	Regulator Outlet Range	Flow Rate
Anaerobic Gas Mix	1/4" BSP male fitting or connection for 6mm Polyurethane tubing. Hydrogen/Anaerobic Gas Mixture Regulator – Two	Two stage	4 - 6 bar (60-90 psi)	40 litres per minute (dynamic)
Nitrogen	Stage – order Code A01745 '/" BSP male fitting or connection for 8mm Polyurethane tubing Nitrogen Regulator – Two Stage – order Code A01748	Two Stage	4 - 6 bar (60-90 psi)	Minimum 100 litres per minute (dynamic)

Notes: If the instrument is to be connected to a single cylinder of anaerobic gas mixture only, then this must be capable of delivering 100 litres/minute.

DWS/SSP0005 Revision Level: 01

Connection Type	Push in connection					
	OR					
	OK .					
	Push on connection					
Mains Requirements						
Electricity Supply	Wall Socket					
240 Volts	1 x Three Pin, 13 Amp. Minimum 6A rating					
Other Considerations Although the workstation should be located in a well ventilated area, avoid close proximity to air conditioning systems and draughts caused by windows and doors. Remember, if you do not have the required regulators you can order them from Don Whitley Scientific:: • Hydrogen/Anaerobic Gas Mixture Regulator – Two Stage – order Code A01745						
	Two Stage – order Code A01748					
Notes If an existing unit is being taken in part exchange or is being removed from the laboratory, it must be de-contaminated before DWS staff handle the unit. A certificate or signed letter confirming the unit has been decontaminated must be given to our engineer.						
In the UK, delivery and installation are free of charge (unless otherwise agreed). If our engineers are unable to install the unit and a return journey is necessary, a charge may be made . Export customers, please refer to your local distributor.						

It is essential that this form is completed and returned, to avoid delay to your installation.

THANK YOU FOR THINKING WHITLEY

Signature	Title	
Print Name	Establishment	

DWS/SSP0005 Revision Level: 01